



SOLICITATION NUMBER: BV3902AD5
DUE DATE: 11/19/2002 @ 3:00 P.M.

STATE PURCHASING AGENT: BRENDA VELDEVERE
DATE ADDENDUM SENT: 10/28/2002

ITEM TO BE PURCHASED: **STATEWIDE CONTRACT FOR ON-SITE DRUG DETECTION
SCREENING TEST KITS**

ADDENDUM #5

PLEASE NOTE THE FOLLOWING:

1. CLARIFICATION OF TERMS IN RFP

FALSE POSITIVE:

A FALSE POSITIVE IS WHEN THE ON-SITE TEST SHOW THE SAMPLE IS POSITIVE FOR A PARTICULAR DRUG METABOLITE AND WHEN SENT TO LAB FOR GC/MS CONFIRMATION IT IS FOUND TO ACTUALLY BE NEGATIVE.

FALSE NEGATIVE:

A FALSE NEGATIVE IS WHEN THE ON-SITE TEST SHOWS NEGATIVE FOR A PARTICULAR DRUG METABOLITE AND IT SENT TO THE LAB FOR GC/MS CONFIRMATION AND IT IS ACTUALLY POSITIVE. THIS IS USUALLY ONLY DONE RANDOMLY AND IN ISOLATED INCIDENTS.

2. THE STATE OF UTAH UNDERSTANDS THE INDUSTRY STANDARD FOR TESTING FOR OPIATES IS 2000 NG/ML BECAUSE OF THE HIGH RISK OF FALSE POSITIVES WHEN USING THE 300 NG/ML CUTOFF LEVEL. THEREFORE, **PLEASE PROVIDE PRICING FOR BOTH THE 300 NG/ML AND THE 2000 NG/ML CUTOFF LEVELS.**
3. THE STATE OF UTAH HAS DETERMINED THAT URINE IS NOT CONSIDERED A RELIABLE METHOD FOR TESTING FOR ALCOHOL. **THEREFORE, THIS RFP WILL ONLY CONSIDER ALCOHOL TESTING KITS THAT USE BREATH OR SALIVA BUT NOT URINE.**

**THIS RFP IS DUE
TUESDAY, NOVEMBER 19, 2002
3:00 P.M.**

To acknowledge receipt of addendum, include a signed and dated copy of this addendum with RFP submittal. It shall be the responsibility of the offeror to appropriately disseminate this information to all concerned prior to the assigned RFP due date.

COMPANY NAME:

NAME:

DATE: